



REMOVABLE ACCESS PANEL TO INTERLOCK MECHANISM

PULL-ACTION TOGGLE CLAMPS TOP & BOTTOM (BOTH SIDES)

VIEWING WINDOW (BOTH SIDES)

S/S FLANGE TO BE APPLIED TO SIDE OF BOX & SEALED AT INSTALLATION

DOOR INTERLOCK

10.375 INSIDE DIMS

26.50

18.00

20.50

24.00

3.00

3.00

6.00 WALL THICKNESS REF ONLY

12.00

AIR TIGHT DOOR SEAL

AIR TIGHT WINDOW SEAL

DETAIL A SCALE 1:1

13.75 INSIDE DIMS

21.500 INSIDE DIMS

14 GA S/S T304, #4 CONSTRUCTION

APPROVED

Print Name:

Date:

Signature:

CONSTRUCTION NOTES

- CUT HOLE THROUGH WALL 24-1/2" HIGH & 18-1/2" WIDE
- INSTALL CABINET THROUGH WALL
- RESEAL WALL OPENING AS REQUIRED
- INSTALL S/S TRIM FLANGES
- CABINET MAY BE INVERTED TO ACCOMODATE DOOR SWING

TBJ
INCORPORATED
1671 Orchard Drive
Chambersburg, PA 17201
Phone 717-261-9700
www.tbjinc.com

DATE: 9/7/2010 | Drawn By - NES | Ref DWG - R10109

TITLE: PT 18-12-24 PASS THRU CABINET

TYPE: ROUGH-IN ASSEMBLY

SIZE	PART NO.	REV.
B	R10152	A

LTR	DESCRIPTION	DATE	BY
	REVISIONS		

NOT TO SCALE | WEIGHT: 60.36 | SHEET 1 OF 1