



INSTALL CYLINDER PLUMB

TBJ INC NOT RESPONSIBLE
FOR STRAINER OF CONCRETE

NOTE:
INCOMING ELECTRICAL CONNECTION (208 V., 3 PH., 60 Hz.) CAN BE LOCATED ANYWHERE ON FLOOR AT SUBLEVEL IN CLOSE PROXIMITY TO THE CYLINDER AND RESERVOIR / POWER UNIT. RESERVOIR / POWER UNIT NOT SHOWN.



SECTION A-A
SCALE: 6"=1'-0"

LET	REVISION	BY	DATE
B	ADDED NOTE	CWF	9-22-08
A	ADDED GASKET & ROD WIPER, INCREASED LOW POSITION HT. 6" - WAS 7-3/8", NOW IT'S 13-5/8"	ALE	4-7-05

TBJ INC
1671 ORCHARD DRIVE
CHAMBERSBURG, PA 17201

SCALE 1-1/2"=1'-0" | DRAWN BY ALE
REF DWG - MSB-4467 | DATE 9-17-03

(LG ANIMAL) NECROPSY TABLE TOP

66-108 HT-C | R03125

DIMENSIONAL TOLERANCE UNLESS OTHERWISE NOTED	
FRACTIONS	±1/64
DECIMAL .XXX	±.005
DECIMAL .XX	±.010
ANGULARITY	±1/2°